PHOTODIODE AMPLIFIER: REFERENCE STANDARD

with traceable calibration





This module is an upgrade for all sglux amplifier boards. It consists of a shielded housing and comes with a traceable calibration and related certificate.

The Module is to be used as a reference Standard for the sglux photodiode amplifier boards or own photodiode amplifier circuit development.

Feature Overview

Measurement properties	Please refer to the related sglux amplifier board.
Input / Output	Input via BNC plugs, output via banana plugs
Housing	Powder-coated aluminium housing with good EMC conditions, rubber feet

Power supply, BNC cable, case

Traceable Calibration

Delivered with

The calibration is performed with equipment directly or indirectly traceable to calibration laboratory reference standards. The reference standards are traceable to the National Institute of Standards and Technology (NIST), SP Technical Research Institute of Sweden, National Physical Laboratory (NPL), Physikalisch-Technische Bundesanstalt (PTB), UKAS Accredited Laboratory 0029 or to the DANAK Accredited Laboratory 333.

Specifications

Parameter	Value	Unit
Degree of protection	IP54	-
Operating temperature	-40+80	${\mathfrak C}$
Storage temperature	-40 +85	${\mathfrak C}$
Power supply	1824	V_{DC}
Power consumption (24V)	10	mA
Weight	0,54	kg

Rev. 1.0 page 1

Manufacturer: **sg**/ux GmbH, Max-Planck-Str. 3, D-12489 Berlin, Tel. +49 30 5301 5211, Fax +49 30 5301 5209 mail: welcome@sglux.de - web: www.sglux.de - WEEE No. DE 76297302

Distributed by: PRC Technologies Corp., Ltd. Tel: 02 530 1714, 02 530 1619, 02 530 1621 Fax: 02 530 1731

Email: info@prctechnologies.co.th, www.prctech-th.com