# **UV Sensor "UV-Surface"**



Top looking surface-mount UV sensor

### GENERAL FEATURES





#### **Properties of this sensor**

The "UV-Surface" sensor is used for UV radiation reference measurements on surfaces exposed to UV light. The sensor contains integrated electronics and is shielded against electro-magnetic interference. Sensor configuration options are spectral response, signal output type and measuring range. The signal output is either a voltage of 0 to 5 V, a current of 4 to 20 mA, CAN bus interface or USB. The UV sensor is available with a NIST or PTB traceable calibration. Cosine correction is available on request.

The measuring range of **analog sglux UV sensors** is 3 orders of magnitude corresponding to 5 mV to 5 V or 4.02 mA to 20 mA output. The highest sensitivity range is 1 nW/cm<sup>2</sup> to 1  $\mu$ W/cm<sup>2</sup>. The lowest sensitivity range is 20 mW/cm<sup>2</sup> to 20 W/cm<sup>2</sup>. The **digital sglux UV sensors** contain an integrated microprocessor that converts the UV radiation into 125kbit/s digital CAN bus data. A large dynamic range of 5 orders of magnitude allows to measure low radiation and strong radiation without changing the probe. Customers may specify any range between the mentioned limits.

Page 3 of this datasheet allows to enter requirements of the needed sensor. After selection you may forward this document to factory or agent, or alternatively use the sensor probe online configurator at www.sglux.com. Please contact us for assistance.

#### **SPECIFICATIONS**

Fixed Specifications Parameter Value

Dimensions please refer to drawing on page 2

Weight 56 g

Temperature Coefficient (30 to 65°C) 0.05 to 0.075%/K

Operating Temperature -20 to +80°C

Storage Temperature -40 to +80°C

Humidity < 80%, non condensing

Configurable Specifications Parameter Value (page 3 shows more detailed information)

Spectral Sensitivity Broadband UV, UVA, UVB, UVC, UV-Index, Bluelight and UV+VIS

Signal Output o to 5 V or 4 to 20 mA or CAN bus signal (125kbit/s) or USB

Current Consumption for o to 5 V = < 30 mA / for 4 to 20 mA = signal out / digital = < 17 mA

Connections cable = 2 m cable with tinned leads on free end

CAN = 2 m cable with 8 pin male connector (to converter or else)

USB = with 1.5 m cable with USB-A plug

Measuring Range between 1 nW/cm $^2$  to 1  $\mu$ W/cm $^2$  and 20 mW/cm $^2$  to 20 W/cm $^2$  for analog

or 100 μW/cm² to 20 W/cm² for digital sensors (see questionaire sheet)




# **UV Sensor "UV-Surface"**



Top looking surface-mount UV sensor


2/2


### FIELD OF VIEW




### DRAWING


#### ANALOG CABLE









#### DIGITAL







M 16 x 0.75

KFV 80 plug




pin layout

# **UV Sensor "UV-Surface"**



Requirements questionaire sheet

### STEP 1 ---- Configuration of Normalized Spectral Responsivity



### STEP 2 ---- Signal Output Type Selection

Please tick your selection. The pin configuration is shown in drawings on page 2.

Output Type	Description	Connection = "cable"
o to 5 V	o to 5 V voltage output proportional to radiation input. Supply voltage is 7 to 24VDC, current consumption is $<$ 30 mA.	
4 to 20 mA	4 to 20 mA current loop for PLC controllers. The current is proportional to the radiation, supply voltage is 24VDC.	$V_{-} = brown,$ $V_{+} = white,$ $shield = black$
CAN bus signal	VSCP protocol according to the following specifications: http://download.sglux.de/probes-digital/vscp-protocol/	Pins 1 & 7 = CAN low Pins 3 & 8 = CAN high Pins 2 & 4 & 5 = GND
USB	The signal is transmitted via standard USB-A plug to a computer. Software and 1.5 m cable are included.	

## STEP 3 ---- Measurement Range Selection

Please mark your approx. max. UV intensity to be measured. The dynamic range for analog UV sensors is 3 orders of magnitude and for digital UV sensors it is 5 orders of magnitude.

max. UV	\\\ / ===3				m-\M / - m-2		. 141/2002	\\\/ 3	a - M/ama2
intensity	1µW/cm²	10μW/cm²	100μW/cm²	1 mw/cm <sup>2</sup>	10mw/cm²	100mvv/cm²	1 W/cm²	10 W/cm <sup>2</sup>	20 W/Cm²



# Sensor Probes Overview and Accessories



### SENSOR PROBES OVERVIEW



**UV-Surface** — Top looking surface-mount UV sensor

For UV radiation reference measurements of radiation exposed to a surface (diameter 38 mm).



UV-Air ..... Threaded body UV sensor

With M22x1.5 thread for many mounting possibilities i.e. inside UV radiation chambers.



**UV-Cosine** — Waterproof cosine corrected UV sensor for outdoor use

Stain repellent for outdoor or in-water measurements. Particularly suited for UV-Index measurements.



**UV-Water-G3/4** •••• 10 bar water pressure proof UV sensor with G3/4" thread

Used in pressurized water systems. Suited for low and medium pressure lamps.



**UV-Water-PTFE** -----> 10 bar water pressure proof UV sensor with G1/4" thread

Used in pressurized water systems. Suited for low pressure lamps.



Complies with standard DVGW294-3(2006), suited for certified water purifiers.



**UV-DVGW-160** — UV sensor for DVGW (160°) and ÖNORM certified water purifiers

Complies with standard DVGW294-3(2006) and ÖNORM 5873-2, suited for certified water purifiers with 160° FOV.



**UV-Cure** — Sensor for strong UV irradiation, working temperature up to 170° (338°F)

To control curing processes or other high temperature operations where strong UV light is present.



**TOCON-Probe** ---- Miniature UV sensor

Miniature UV sensor in M12x1 housing. Available with o to 5 V voltage output.

# ACCESSORIES FOR ANALOG SENSOR PROBES



**Sensor Monitor 5.0** measuring and control module



# ACCESSORIES FOR DIGITAL SENSOR PROBES





**DIGIBOX** ---->
CAN-to-USB converter



Control Pad 
windows 8 based 10.1"
tablet computer
display unit

#### WINDOWS



win294 ····
measurement window
acc. to DVGW 294-3
and ÖNORM M5873

